



校准证书
CALIBRATION CERTIFICATE



证书编号:

Certificate No.



扫一扫验真伪

2WB23031752543-0006

委托方:

Client

陕西高研智能科技有限公司

委托方地址:

Address

西安市高新区西沣南路1588号

仪器/样品名称:

Description

SF6气体检漏仪

型号/规格:

Model/Type

HM3020X

制造厂商:

Manufacturer

武汉国电华美电气设备有限公司

出厂编号:

Serial No.

0220910

管理号:

Asset No.

样品接收日期:

Date of Receipt

2023-03-22

Y M D

结果:

Conclusion

所校准项目合格(Passed at Calibration Items)

校准日期:

Date of Calibration

2023-03-22

Y M D

建议下次校准日期:

Due Date

2024-03-21

Y M D

校准:

Calibration by

程越

审核:

Inspected by

徐继胜

授权签字人:

Approved Signatory

方文潮

证书专用章

(Stamp)

本实验室地址: 广东省广州市番禺区石碁镇农科所南街8号 @力赛计量实验室

Address: No.8.South Street Shi Ji Institute Guangzhou.Guangdong.China

联系电话(Tel): 020-31134076

传真(Fax): 020-31134076

公司网址(Web): www.lisaitest.com

投诉电话(CT): 020-31104772

邮政编码(Post): 511400

电子邮件(E-mail): cal@lisaitest.com



校准说明

DIRECTIONS OF CALIBRATION

证书编号:
Certificate No.



2WB23031752543-0006

本机构质量管理体系符合ISO/IEC17025标准要求。

The quality system is in accordance with ISO/IEC17025.

1. 证书内页中"P"代表"合格";"F"代表"不合格";"N/A"代表"不适用"。

In the datasheet,"P" represents "Pass" and "F" represents "Fail" and "N/A" represents "Not Applicable".

2. 本证书编号具有唯一性, 后缀若带有"A~Z"的证书为替换证书, 自发出后原证书即刻作废。

Each certificate has a unique number,The suffix of "A~Z" will be added to the number as a replacement of the old version. The original certificate will be officially invalid once the new certificate number is issued.

3. 环境条件。(Environmental condition during the calibration).

温度(Temperature): 22°C 相对湿度 (Relative Humidity): 58 %

4. 校准地点。(Place of the Calibration).

客户现场(1楼车间)

5. 未加盖“证书报告专用章”无效,报告无校准或检测、审核、授权签字人签章无效。

The report is invalid without the official stamp, report is invalid without 'report stamp', The report is invalid without the signatures of Approval and Reviewer.

6. 样品及生产/研制方的相关信息由委托方提供并由委托方承担相应责任。

The information of the samples and the manufacturer/developer shall be provided by the applicants and applicant shall bear the corresponding responsibilities.

7. 本次校准的技术依据及CNAS认可范围, 超出范围的内容未被认可。详细认可范围请查看CNAS网站中注册编号为L7127的证书附件。(Reference documents and accredited scope by CNAS for the Calibration beyond which isn't accredited. Please see the attachment of certificate No.L7127 on CNAS website for details.)

JJF 1263-2010 六氟化硫检测报警仪校准规范 Calibration Specification for the Alarmer Detector of Sulfur Hexafluoride

8. 本次校准使用的主要计量标准器具。(Main standards of measurement used in the Calibration).

| 名称 Description | 出厂编号 Serial No. | 证书号/有效期 Certificate No./ Due Date | 溯源机构 Traceability Institute | 技术特征 Technique Character |
|-------------------|--------------------|--------------------------------------|--------------------------------|-----------------------------|
| 空气中六氟化硫气体标准物质 | 1911102130 | GBW(E)061075 / 2023-03-24 | 南京长元工业气体有限公司 | 二级标物 |
| 空气中六氟化硫气体标准物质 | 61907164 | GBW(E)061075 / 2023-03-24 | 南京长元工业气体有限公司 | 二级标物 |



广州力赛计量检测有限公司

GUANGZHOU LISAI CALIBRATION AND TESTING CO.,LTD

证书编号:
Certificate No.



中国认可
国际互认
校准
CALIBRATION
CNAS L7127

2WB23031752543-0006

1、外观以及一般性检查: 正常
In view of External and Generality check: Pass

2、报警功能运作状况 (Alarm function) :

标准气体: 六氟化硫
The standard gas: SF₆

| 标气浓度 The gas concentra (ppm) | 报警功能运作 Alarm function | 结论 Conclusion (Pass/Fail) |
|------------------------------------|--------------------------|---------------------------------|
| 100 | 正常 | P |
| 1000 | 正常 | P |

附: 关于测量结果不确定度的说明:
appendix: Directions of uncertainty in the calibration

1.依据(Reference document)

JJF 1059.1-2012 测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

2.本报告给出的扩展不确定度是由合成标准不确定度乘以包含概率约为95%时对应的包含因子k得到的。

The expanded uncertainty given in this report is obtained by multiplying the combined standard uncertainty by the corresponding coverage factor k when the coverage probability of about 95%.

-----以下空白(The below is blank)-----



广州力赛计量检测有限公司

GUANGZHOU LISAI CALIBRATION AND TESTING CO.,LTD



中国认可
国际互认
校准
CALIBRATION
CNAS L7127

证书编号:
Certificate No.



2WB23031752543-0006

本页空白